

30GL Slit Lamp Microscope Specification

Model name		30GL Slit Lamp Microscope
Microscope	Type	Greenough binocular stereo microscope
	Magnification changer	Revolver objective changing system with 2 magnification changes
	Eyepiece	10x / 16x
	Total magnifications	10x / 16x (When 10x eyepieces are in use) 16x / 25x (When 16x eyepieces are in use)
	Field of view (diameter)	18 / 11.25 mm (When 10x eyepieces are in use) 15 / 9.3 mm (When 16x eyepieces are in use)
	Stereo angle	13°
	Interpupillary distance	52-82 mm
	Eyepiece diopter range	± 8 D (When 10x eyepieces are in use) ± 10 D (When 16x eyepieces are in use)
	Working distance	94 mm (from objective lens to patient's eye)
Illumination Unit	Slit image width	Continuously variable 0-14 mm (almost circular at 14 mm)
	Slit image length	Continuously variable 1-14 mm
	Slit image rotation	Continuously variable 0-180° with 3 click stops at 45°, 90°, 135°
	Slit decentration (horizontal)	Possible (fixable at 0°)
	Slit inclination (vertical)	0° / 5° / 10° / 15° / 20°
	Illumination field diameter	14 / 10 / 5 / 3 / 2 / 1 / 0.2 mm
	Scanning (decentration) angle	7° to the left, 7° to the right
	Working distance (from mirror/prism to patient's eye)	80 mm
	Filter	Blue / Green (red-free) / Blue correction / Blue cut
Base Unit	Light source	LED (3500K)
	Base unit operation	Single-handed joystick operation for forward-backward, left-right, up-down movement.
	Movement range	105 mm (forward-backward, left-right), 30 mm (up-down)
Headrest Unit	Fine movement range	15 mm (forward-backward, left-right)
	Vertical movement range	95 mm
Power Unit	Fixation light	LED (red)
	Power supply	AC 100-240 V (±10%), 0.4-0.2 A
	Power input	DC 5 V 2 A
Weight (excl. power adapter & options)		12.5 kg

*Specifications as detailed in this sheet may be subject to change as improvements are made to products.

*For the latest information, please visit our website (<https://www.takagi-j.com/en/>).