Disinfectant list for S12-18 Measuring Prism

The following is a list of disinfectants that can be used in addition to the disinfectants listed in the instruction manual. The concentration does not indicate the bactericidal effect of the disinfectant, but indicates the range in which the prism body can withstand each disinfectant.



Do not use disinfectants containing alcohol as they will damage the prism body.
Please note that the warranty does not cover damage to the prism body using disinfectants not listed in the instruction manual or this disinfectant list.

Disinfectant	Concentration	Maximum allowable number of disinfections	Maximum penetration length / time
NaOCl (Sodium hypochlorite)	0.05%	1,600 times	10 minutes
NaOCl (Sodium hypochlorite)	0.5%	1,600 times	10 minutes
H_2O_2 (Hydrogen peroxide)	3%	1,600 times	10 minutes
Cidex®plus	3.5%	1,600 times	10 minutes
Hibitane®	0.5%	1,600 times	10 minutes
Sekusept®Activ	2%	1,600 times	15 minutes
Sekusept®PLUS	4%	1,600 times	15 minutes
Sekusept®Extra N	2%	1,600 times	15 minutes

www.takagi-j.com

For the Americas, Asia-Pacific & Middle East

Takagi Seiko Co., Ltd.

330-2 Iwafune
Nakano-shi
Nagano-ken
383-8585
Japan

+81 (0)269-22-4511

www.takagieurope.com

For Europe & Africa

Takagi Ophthalmic Instruments Europe Ltd

Citylabs 1.0
Nelson Street
Manchester
M13 9NQ
United Kingdom

+44 (0)161 273 6330

©2026 Takagi Seiko Co., Ltd. All rights reserved.

J(S12-18DL)EN Rev.0

TAKAGI

Serving Your Vision